



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

SIGMA SCIENTIFIC SERVICES LLC  
490 Sawgrass Corporate Parkway Suite 130  
Sunrise, FL 33325  
Mr. Jorge Millan E-Mail: [jmillan@sigmabiomedical.com](mailto:jmillan@sigmabiomedical.com)  
Web: [www.sigmascientific.us](http://www.sigmascientific.us)

ELECTRICAL

Valid To: December 31, 2027

Certificate Number: 4896.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's R256 - Specific Requirements<sup>1</sup>), accreditation is granted to this laboratory to perform the following electrical tests:

<u>Test Technology:</u>	<u>Test Method(s):</u>
Measurement and Characterization of Medical Ultrasonic Fields	IEC 62127-1 Edition 1.1 2013; IEC 62359 Edition 2.1 2017
Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment	AIUM NEMA UD2
Medical Electrical Equipment – Part 2-37: Particular Requirements for the Basic Safety and Essential Performance of Ultrasonic Medical Diagnostic and Monitoring Equipment	IEC 60601-2-37:2024
Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance	IEC 60601-1 Edition 3.2 2020-08-Clauses 4-9, 11-16; ANSI AAMI ES60601-1:2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012 <sup>3</sup>
Medical Electrical Equipment- Part 2-5: Particular Requirements for the Basic Safety and Essential Performance of Ultrasonic Physiotherapy Equipment	IEC 60601-2-5 Edition 3.0 2009-07
Medical Electrical Equipment- Part 2-62: Particular Requirements for the Basic Safety and Essential Performance of High Intensity Therapeutic Ultrasound (HITU) Equipment	IEC 60601-2-62 Edition 1.0 2013-07

<b><u>Test Technology:</u></b>	<b><u>Test Method(s):</u></b>
Medical Electrical Equipment- Part 1-11: General Requirements for the Basic Safety and Essential Performance- Collateral Standard: Requirements for Medical Electrical Equipment and Medical Electrical Systems Used in the Home Healthcare Environment	IEC 60601-1-11 Edition 2.1 2020-07
General requirements for basic safety and essential performance- Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	IEC/EN 60601-1-8 (Ed. 2 + Am 1 + Am 2); ANSI AAMI IEC 60601-1-8:2006 and A1:2012
General methods of measuring the performance of ultrasonic pulsed-echo diagnostic equipment	JIS T 1501: 2005
Ultrasonics –Characteristic of Fields	IEC 61846 Edition 1998-04; IEC 63045 Edition 2020-05; IEC 61689 Edition 3.0 2013-02
Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability	IEC 60601-1-6 Edition 3.2 2020-07
<b><u>Infant Incubators</u></b>	
Medical electrical equipment - Part 2-19: Particular requirements for the basic safety and essential performance of infant incubators	IEC 60601-2-19 Edition 3.0 2020-09
Medical electrical equipment - Part 2-20: Particular requirements for the basic safety and essential performance of infant transport incubators	IEC 60601-2-20 Edition 3.0 2020-09
<b><u>Medical Endoscopes</u></b>	
Optics and photonics — Medical endoscopes and Endotherapy devices —Part 1: General requirements	ISO 8600-1 2 <sup>nd</sup> Edition (2005-05-1)
Determination of field of view and direction of view of endoscopes with optics	ISO 8600-3 2 <sup>nd</sup> Edition (2019-8)
Determination of maximum width of insertion portion	ISO 8600-4 3 <sup>rd</sup> Edition (2023-01)
Determination of optical resolution of rigid endoscopes with optics	ISO 8600-5 2 <sup>nd</sup> Edition (2020-10)

<b><u>Test Technology:</u></b>	<b><u>Test Method(s):</u></b>
<b><u>Medical Endoscopes (continued)</u></b>	
Photography — Electronic still-picture imaging — Noise measurements	ISO 15739 3 <sup>rd</sup> Edition (2017-05)
Photography — Electronic still picture imaging — Resolution and spatial frequency responses	ISO 12233 4 <sup>th</sup> Edition (2023-02)
Medical electrical equipment – Part 2-18: Particular requirements for the basic safety and essential performance of endoscopic equipment	IEC 60601-2-18 Edition 3.0 2009-08
Photobiological safety of lamps and lamp systems	IEC 62471 Edition 1.0 b:2006 Exclusion: Wavelength range 1100nm - 3000nm
<b><u>Ophthalmic Instruments</u></b>	
Ophthalmic instruments – Fundamental Requirements and Test Methods-Part 1: General Requirements Applicable to All Ophthalmic Instruments	ISO 15004-1:2020 Exclusion: Wavelength range 1150nm - 2500nm
Ophthalmic instruments – Fundamental requirements and Test Methods-Part 2: Light Hazard Protection	ISO 15004-2:2007 Exclusion: Wavelength range 1150nm - 2500nm

**On the following products and materials:**

Diagnostic Ultrasound Systems and Transducers  
 Ultrasonics – Physiotherapy Systems  
 Medical Equipment and Medical Devices  
 High Intensity Therapeutic Ultrasound (HITU) Equipment

<sup>1</sup>*The laboratory is only accredited for testing activities outlined within the test methods listed above. Reference to any other activity within these standards, such as risk management or risk assessment, does not fall within the laboratory's accredited capabilities.*

<sup>2</sup>*The following exclusions apply to the ANSI/AAMI ES60601: Clauses 10.1-10.7.*

*Note: Exclusions listed for ANSI AAMI ES60601-1 also apply to other collaterals and particulars in the 60601/80601 family of standards except where otherwise stated.*



## Accredited Laboratory

A2LA has accredited

### SIGMA SCIENTIFIC SERVICES LLC

*Sunrise, FL*

for technical competence in the field of

### Electrical Testing

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 15<sup>th</sup> day of April 2026.

A blue ink signature of Mr. Trace McInturff, written in a cursive style.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 4896.01  
Valid to December 31, 2027

*For the tests to which this accreditation applies, please refer to the laboratory's Electrical Scope of Accreditation.*